



**ADDENDUM  
PURCHASING OFFICE  
CITY OF AUSTIN, TEXAS**

---

**Solicitation: RFP 9200 ELF3002**

**Addendum No: 2**

**Date of Addendum: 06/06/2019**

---

This addendum is to incorporate the following changes, questions and answers to the above referenced solicitation.

**I. Questions and Answers:**

**I.1 (a) Question:**

To clarify, are you in need of a digital (DR) upgrade for an existing veterinary x-ray system (i.e. flat panel detector retrofit kit with computer workstation and software) or an entirely new veterinary DR suite (i.e. table with tubestand plus DR detector and computer workstation with software)?

**Answer:**

The City would prefer to replace the workstation and software only due to cost. If replacing non-computer equipment is essential for long term compatibility, the City will definitely consider it.

If replacing the table, tubestand, and computer workstation and software is required the City can replace it. The City believes the existing table will work with any system and do not think it will need to be replaced. The City assumes the detector would be replaced/upgraded with the computer workstation and software.

**(b) Question:**

Requirement 2.19 seemed to imply you needed a DR suite, but with the brand name included we wanted to clarify that was a request for new equipment rather than a listing of specifications on existing equipment to be upgraded.

**Answer:**

The City included a list of our current equipment to determine compatibility with a new computer workstation and software suite. The City only wants to replace it if necessary.

**I.2 Question:**

Unfortunately we were unable to attend the preproposal conference call earlier today. Will answers to other vendor questions that were provided on the call be posted via RFP addendum/amendment?

**Answer:**

Per Addendum 1 of Solicitation, the Pre-Proposal Meeting Date, Time and Teleconference information changed from June 3<sup>rd</sup>, 2019 to June 4<sup>th</sup>, 2019.

New Date, Time and Teleconference Information:

June 4, 2019 @ 8:30 am CST (Local Time)  
Teleconference Number: 512-974-9300  
Participant Code: 851602

Verbal questions and answers discussed during the Pre-Proposal meeting will not be posted via an Addendum.

Revised 12/13/2015

All questions during the remainder of the solicitation are required to be submitted in writing to Elisa Folco, authorized point of contact for this solicitation. All questions and answers will be posted via an Addendum.

II. ALL OTHER TERMS AND CONDITIONS REMAIN THE SAME.

APPROVED BY: Elisa Folco  
Elisa Folco, Procurement Specialist IV  
Purchasing Office, 512-974-1421

06/06/19  
Date